










## Running machine h/p/cosmos quasar med 4.0 - Specifications



[ Picture with standard handrail and user terminal ]

-  [Operation](#)
-  [Standards / options / accessories](#)
-  [quasar med vs. mercury med](#)
-  [Specifications profile form](#)
-  [Printout \[ \\*.pdf \]](#)
-  [Specifications \[ \\*.txt \]](#)
-  [Gallery \[ \\*.gif / \\*.jpg / \\*.eps \]](#)
-  [Model overview](#)

Speed	0 ... 25 km/h / ... 6.9 m/s / ... 15.5 mph	medical	Running surface	170 x 65 cm / 66.0 x 25.6 inches
Elevation	0 ... 28 % / ... 15.6°		Track access height **	23 cm / 9.0 inches
Drive motor	2.2 kW / 3 hp		Device weight **	380 kg / 840 lbs
Power supply Fuse	220 ... 240 V / 15 A optionally special voltage		Frame size **	230 x 105 x 140 cm 90.6 x 41.3 x 55.1 inches
Leakage current	0.2 mA *		max. load at surface	200 kg / 442 lbs
* Maximum leakage current in accordance to provision IEC EN 60601-1: 0.5 mA. ** Data of dimension and weight of the device can deviate.			Live load p **	7.11 kN/m <sup>2</sup> / 725 kg/m <sup>2</sup> - dynamic (subject's weight 200 kg plus factor dynamic and safety coefficient)
h/p/cosmos running machines of the category medical meet the following technical safety standards and legal demands for medical application: MDD class <b>IIb</b> , <b>CE 0123</b> , IEC EN 60601-1, IEC EN 60601-1-2, IEC EN 60601-1-4, EN 957-1, EN 957-6 and VDE 0750/0751.		 According to EU-provisions it is prohibited to place on market sports running machines in the medical field.		

order number	
h/p/cosmos quasar med 4.0	cos03001100000